



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

<b>Service</b>
<b>Report</b>

<b>RMA Number</b>	89690
-------------------	-------

### Customer Information:

<b>Company</b>	WEBB RESEARCH CORPORATION	<b>Date</b>	6/30/2016
<b>Contact</b>	CHARLES STILL		
<b>PO Number</b>	PW05088		

<b>Serial Number</b>	SLOCUM-9086
<b>Model Number</b>	SLOCUM

### Services Requested:

1. Evaluate/Repair Instrumentation.
2. Perform Routine Calibration Service.

### Problems Found:

1. bad endcap. Anodizing in poor condition and damaged screw hole.

### Services Performed:

1. Performed initial diagnostic evaluation.
2. Calibrated the pressure sensor.
3. Performed "Post Cruise" calibration of the temperature & conductivity sensors.
4. WEBB PAYLOAD CTD, ENDCAP, HARDCOATED
5. Performed complete system check and full diagnostic evaluation.

### Special Notes:

--

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9086  
CALIBRATION DATE: 21-Jun-16

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter


**COEFFICIENTS:**

g = -9.812300e-001  
h = 1.373538e-001  
i = -3.060588e-004  
j = 4.070468e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 1.1824e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2677.91	0.00000	0.00000
1.0000	34.6915	2.96631	5370.16	2.96631	-0.00000
4.5000	34.6717	3.27242	5573.92	3.27242	-0.00000
15.0000	34.6294	4.25110	6179.68	4.25112	0.00001
18.5000	34.6208	4.59522	6378.72	4.59522	-0.00000
24.0000	34.6115	5.15152	6687.64	5.15151	-0.00001
29.0000	34.6070	5.67188	6963.88	5.67188	0.00000
32.4999	34.6052	6.04332	7154.32	6.04332	0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$   
 $t = \text{temperature (°C)}$ ;  $p = \text{pressure (decibars)}$ ;  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;  
 $\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$   
 $\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9086  
CALIBRATION DATE: 21-Jun-16

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.080080e-004  
a1 = 3.093869e-004  
a2 = -4.762222e-006  
a3 = 2.092634e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	578912.5	1.0000	0.0000
4.5000	494828.2	4.5000	-0.0000
15.0000	315156.3	15.0000	0.0000
18.5000	272858.0	18.5000	-0.0000
24.0000	218870.4	24.0000	-0.0000
29.0000	180221.2	29.0000	0.0000
32.4999	157832.9	32.4999	-0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) =  $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9086  
CALIBRATION DATE: 16-Jun-16

Slocum Payload CTD PRESSURE CALIBRATION DATA  
1450 psia S/N 3846545

**COEFFICIENTS:**

PA0 =	-3.057858e-001	PTCA0 =	5.237789e+005
PA1 =	4.592741e-003	PTCA1 =	-3.856183e+000
PA2 =	-1.860757e-011	PTCA2 =	3.913964e-002
PTEMPA0 =	-7.712954e+001	PTCB0 =	2.536613e+001
PTEMPA1 =	5.087796e-002	PTCB1 =	-1.750000e-004
PTEMPA2 =	-6.304203e-007	PTCB2 =	0.000000e+000

**PRESSURE SPAN CALIBRATION**

**THERMAL CORRECTION**

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.64	526966.0	2017.0	14.65	0.00	32.50	2216	527012.20
315.00	592681.0	2027.0	316.42	0.10	29.00	2143	527016.40
614.88	657713.0	2028.0	614.90	0.00	24.00	2039	527023.80
914.88	723098.0	2030.0	914.84	-0.00	18.50	1926	527038.60
1215.09	788553.0	2030.0	1214.94	-0.01	15.00	1853	527046.10
1464.88	843097.0	2030.0	1464.90	0.00	4.50	1638	527080.50
1214.92	788573.0	2030.0	1215.03	0.01	1.00	1566	527090.40
914.93	723127.0	2030.0	914.97	0.00			
614.92	657719.0	2030.0	614.93	0.00			
314.93	592349.0	2029.0	314.90	-0.00			
14.64	526965.0	2031.0	14.65	0.00			

TEMPERATURE (°C)	SPAN (mV)
-5.00	25.37
35.00	25.36

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 16-Jun-16 -0.00

