

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9353
 CALIBRATION DATE: 11-Mar-16


Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.852893e-001	CPcor = -9.5700e-008
h = 1.308156e-001	CTcor = 3.2500e-006
i = -1.272570e-004	WBOTC = 9.1032e-008
j = 2.686007e-005	

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2745.97	0.00000	0.00000
1.0000	34.7105	2.96778	5494.81	2.96778	0.00000
4.5000	34.6907	3.27404	5702.93	3.27404	0.00000
15.0000	34.6483	4.25318	6321.71	4.25317	-0.00001
18.5000	34.6394	4.59743	6525.05	4.59742	-0.00000
24.0000	34.6296	5.15391	6840.69	5.15393	0.00001
29.0000	34.6245	5.67443	7122.93	5.67442	-0.00001
32.5000	34.6222	6.04596	7317.49	6.04584	-0.00012

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0
 t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;
 Conductivity (S/m) = (g + h * f² + i * f³ + j * f⁴) / 10 (1 + δ * t + ε * p)
 Residual (Siemens/meter) = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9353
 CALIBRATION DATE: 12-Feb-16

Slocum Payload CTD PRESSURE CALIBRATION DATA
 1450 psia S/N 4449854

COEFFICIENTS:

PA0 =	1.187379e-001	PTCA0 =	5.244773e+005
PA1 =	4.589555e-003	PTCA1 =	3.750621e+000
PA2 =	-1.996881e-011	PTCA2 =	-1.154652e-001
PTEMPA0 =	1.495120e+002	PTCB0 =	2.510188e+001
PTEMPA1 =	-6.161058e-002	PTCB1 =	-2.500000e-005
PTEMPA2 =	-1.297175e-007	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.66	527673.0	2051.0	14.67	0.00	32.50	1892	527658.50
314.89	593083.0	2046.0	314.79	-0.01	29.00	1948	527731.75
614.99	658514.0	2045.0	614.83	-0.01	24.00	2029	527694.25
914.96	723968.0	2044.0	914.80	-0.01	18.50	2117	527705.25
1214.94	789498.0	2041.0	1214.96	0.00	15.00	2173	527712.50
1464.47	843988.0	2041.0	1464.41	-0.00	4.50	2342	527692.50
1214.83	789493.0	2040.0	1214.94	0.01	1.00	2399	527689.25
914.86	724002.0	2041.0	914.96	0.01			
614.77	658535.0	2045.0	614.92	0.01			
314.80	593105.0	2044.0	314.89	0.01			
14.66	527675.0	2046.0	14.68	0.00			

	TEMPERATURE (°C)	SPAN (mV)
	-5.00	25.10
	35.00	25.10

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$


$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 12-Feb-16 -0.00



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9353
CALIBRATION DATE: 11-Mar-16

Slocum Payload CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.371503e-004
a1 = 3.130767e-004
a2 = -4.911443e-006
a3 = 2.123671e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	577351.0	1.0000	-0.0000
4.5000	493850.0	4.5000	0.0000
15.0000	315208.5	14.9999	-0.0001
18.5000	273090.8	18.5001	0.0001
24.0000	219296.5	24.0001	0.0001
29.0000	180750.8	28.9998	-0.0002
32.5000	158401.3	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

Date, Offset (mdeg C)

● 11-Mar-16 0.00

