

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA  
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9066  
 CALIBRATION DATE: 11-Apr-13

SLOCUM PAYLOAD CTD  
 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE


**ITS-90 COEFFICIENTS**

a0 = -9.081145e-005  
 a1 = 3.036532e-004  
 a2 = -4.262464e-006  
 a3 = 1.960693e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	582300.4	1.0001	0.0001
4.5000	497889.2	4.4999	-0.0001
15.0000	317406.4	15.0001	0.0001
18.5000	274890.4	18.5001	0.0001
24.0001	220606.8	23.9999	-0.0002
29.0000	181726.4	29.0000	0.0000
32.5000	159195.2	32.5000	0.0000

Temperature ITS-90 =  $1 / \{ a_0 + a_1 [\ln(n)] + a_2 [\ln^2(n)] + a_3 [\ln^3(n)] \} - 273.15$  (°C)

Residual = instrument temperature - bath temperature



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9066  
 CALIBRATION DATE: 11-Apr-13

SLOCUM PAYLOAD CTD  
 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -9.758777e-001	CPcor = -9.5700e-008
h = 1.509455e-001	CTcor = 3.2500e-006
i = -4.214149e-004	WBCTC = 1.5039e-006
j = 3.486791e-005	


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2548.68	0.00000	0.00000
1.0000	34.8378	2.97763	5129.99	2.97763	0.00001
4.5000	34.8180	3.28487	5325.08	3.28486	-0.00000
15.0000	34.7753	4.26712	5904.89	4.26711	-0.00001
18.5000	34.7660	4.61242	6095.35	4.61242	0.00000
24.0001	34.7557	5.17062	6390.90	5.17063	0.00001
29.0000	34.7494	5.69259	6655.03	5.69259	-0.00001

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = instrument conductivity - bath conductivity



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9066  
 CALIBRATION DATE: 08-Apr-13

SLOCUM PAYLOAD CTD  
 PRESSURE CALIBRATION DATA  
 1450 psia S/N 3819475

**COEFFICIENTS:**

PA0 = 3.454005e-002  
 PA1 = 4.600039e-003  
 PA2 = -1.535458e-011  
 PTEMPA0 = -7.462555e+001  
 PTEMPA1 = 4.983728e-002  
 PTEMPA2 = -3.781723e-007

PTCA0 = 5.245939e+005  
 PTCA1 = -2.732939e+000  
 PTCA2 = 1.400702e-001  
 PTCB0 = 2.536362e+001  
 PTCB1 = -6.750000e-004  
 PTCB2 = 0.000000e+000

**PRESSURE SPAN CALIBRATION**

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FSR
14.66	527784.0	1962.0	14.69	0.00
314.93	593022.0	1963.0	314.89	-0.00
614.93	658246.0	1964.0	614.89	-0.00
914.94	723507.0	1965.0	914.93	-0.00
1214.94	788787.0	1965.0	1214.93	-0.00
1464.99	843212.0	1965.0	1464.95	-0.00
1214.91	788794.0	1965.0	1214.97	0.00
914.88	723510.0	1965.0	914.95	0.00
614.93	658256.0	1965.0	614.94	0.00
314.95	593031.0	1964.0	314.93	-0.00
14.66	527782.0	1965.0	14.68	0.00

**THERMAL CORRECTION**

TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
32.50	2186	527886.00
29.00	2113	527873.60
24.00	2010	527851.20
18.50	1896	527828.40
15.00	1824	527816.40
4.50	1607	527822.60
1.00	1535	527829.40

TEMP (ITS90)	SPAN (mV)
5.00	25.37
35.00	25.34

$$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$$

Date, Avg Delta P %FS

08-Apr-13 0.00

