

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9026
CALIBRATION DATE: 26-Aug-12

SLOCUM PAYLOAD CTD
TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE


ITS-90 COEFFICIENTS

a0 = -7.154871e-005
a1 = 3.012319e-004
a2 = -4.070747e-006
a3 = 1.903265e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	567674.0	1.0001	0.0001
4.5000	485244.3	4.4998	-0.0002
15.0000	309100.2	15.0000	0.0000
18.5000	267629.6	18.5002	0.0002
24.0000	214702.5	23.9999	-0.0001
29.0000	176808.0	28.9998	-0.0002
32.5000	154852.2	32.5001	0.0001

Temperature ITS-90 = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9026
 CALIBRATION DATE: 26-Aug-12

SLOCUM PAYLOAD CTD
 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.846832e-001
 h = 1.328303e-001
 i = -1.339382e-004
 j = 2.688635e-005

CPcor = -9.5700e-008
 CTcor = 3.2500e-006
 WBOTC = 1.5292e-007

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREO (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2724.39	0.00000	0.00000
1.0000	34.9189	2.98390	5464.54	2.98390	0.00001
4.5000	34.8996	3.29181	5671.86	3.29180	-0.00000
15.0000	34.8574	4.27612	6288.17	4.27611	-0.00001
18.5000	34.8482	4.62214	6490.68	4.62214	-0.00000
24.0000	34.8375	5.18143	6804.97	5.18144	0.00001
29.0000	34.8294	5.70422	7085.83	5.70422	-0.00001

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$


$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

Date, Slope Correction

● 26-Aug-12 1.0000000



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9026
 CALIBRATION DATE: 23-Aug-12

SLOCUM PAYLOAD CTD
 PRESSURE CALIBRATION DATA
 1450 psia S/N 3653217

COEFFICIENTS:

PA0 = -7.071768e-003
 PA1 = 4.559772e-003
 PA2 = -2.014527e-011
 PTEMPA0 = -7.094159e+001
 PTEMPA1 = 5.113391e-002
 PTEMPA2 = -4.895939e-007

PTCA0 = 5.246603e+005
 PTCA1 = -3.236157e-001
 PTCA2 = 8.776718e-003
 PTCB0 = 2.534362e+001
 PTCB1 = -2.750000e-004
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FSR
14.65	527883.0	1831.0	14.70	0.00
315.11	593752.0	1834.0	315.02	-0.01
615.07	659589.0	1837.0	615.03	-0.00
915.00	725463.0	1837.0	915.02	0.00
1214.95	791359.0	1839.0	1214.94	-0.00
1464.94	846305.0	1839.0	1464.89	-0.00
1214.91	791362.0	1839.0	1214.95	0.00
914.97	725466.0	1840.0	915.04	0.00
615.03	659589.0	1840.0	615.03	-0.00
315.04	593752.0	1840.0	315.03	-0.00
14.65	527872.0	1841.0	14.65	0.00

THERMAL CORRECTION

TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
32.50	2064	527904.90
29.00	1993	527906.60
24.00	1891	527904.90
18.50	1779	527903.70
15.00	1709	527902.80
4.50	1497	527908.30
1.00	1426	527905.40

TEMP (ITS90)	SPAN (mV)
-5.00	25.34
35.00	25.33

$$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$$

Date, Avg Delta P %FS

● 23-Aug-12 0.00

