

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0104
 CALIBRATION DATE: 11-Dec-11

SLOCUM PAYLOAD CTD
 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = -7.954768e-005
 a1 = 3.055388e-004
 a2 = -4.488936e-006
 a3 = 2.001958e-007


BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	570388.0	1.0000	0.0000
4.5000	487156.0	4.5000	-0.0000
15.0001	309578.8	15.0000	-0.0001
18.5002	267837.0	18.5004	0.0002
24.0000	214620.8	24.0000	-0.0000
29.0000	176562.2	28.9998	-0.0002
32.5000	154530.2	32.5001	0.0001

$$\text{Temperature ITS-90} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15 \text{ (}^\circ\text{C)}$$

$$\text{Residual} = \text{instrument temperature} - \text{bath temperature}$$

Date, Delta T (mdeg C)

11-Dec-11 0.00



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0104
 CALIBRATION DATE: 11-Dec-11

SLOCUM PAYLOAD CTD
 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.843122e-001
 h = 1.465982e-001
 i = -4.133758e-004
 j = 5.174089e-005

CPcor = -9.5700e-008
 CTcor = 3.2500e-006
 WBOTC = 3.8494e-007

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2597.63	0.00000	0.00000
1.0000	34.8264	2.97675	5211.39	2.97674	-0.00001
4.5000	34.8059	3.28384	5409.17	3.28385	0.00001
15.0001	34.7625	4.26572	5997.11	4.26572	0.00000
18.5002	34.7533	4.61093	6190.28	4.61094	0.00001
24.0000	34.7433	5.16897	6490.05	5.16895	-0.00001
29.0000	34.7373	5.69083	6758.02	5.69083	-0.00000
32.5000	34.7337	6.06321	6942.71	6.06322	0.00000

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$


$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

Date, Slope Correction

11-Dec-11 1.0000000



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0104
 CALIBRATION DATE: 09-Dec-11

SLOCUM PAYLOAD CTD
 PRESSURE CALIBRATION DATA
 1450 psia S/N 3459008

COEFFICIENTS:

PA0 = 1.165921e-001	PTCA0 = 5.248777e+005
PA1 = 4.929230e-003	PTCA1 = 2.803011e+000
PA2 = -2.534691e-011	PTCA2 = -7.331970e-002
PTEMPA0 = -7.510170e+001	PTCB0 = 2.531113e+001
PTEMPA1 = 4.835001e-002	PTCB1 = 4.250000e-004
PTEMPA2 = -3.428701e-007	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FSR
14.75	527882.0	2001.0	14.79	0.00
315.06	588831.0	2007.0	315.01	-0.00
615.08	649773.0	2007.0	615.02	-0.00
915.08	710757.0	2008.0	915.04	-0.00
1215.11	771790.0	2010.0	1215.11	-0.00
1465.09	822657.0	2012.0	1465.05	-0.00
1215.07	771796.0	2011.0	1215.13	0.00
915.03	710769.0	2012.0	915.09	0.00
615.03	649781.0	2011.0	615.05	0.00
315.03	588838.0	2012.0	315.05	0.00
14.75	527872.0	2015.0	14.74	-0.00

THERMAL CORRECTION

TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
32.50	2262	527898.20
29.00	2187	527909.60
24.00	2080	527916.00
18.50	1963	527914.00
15.00	1889	527909.60
4.50	1666	527898.80
1.00	1592	527891.00

TEMP (ITS90)	SPAN (mV)
-5.00	25.31
35.00	25.33

$$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$$

Date, Avg Delta P %FS

● 09-Dec-11 -0.00

