

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9288  
CALIBRATION DATE: 26-Feb-15

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## COEFFICIENTS:


a0 = -1.636469e-004  
a1 = 3.168003e-004  
a2 = -5.135613e-006  
a3 = 2.198139e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	575727.0	1.0000	0.0000
4.5000	492898.0	4.5000	-0.0000
15.0000	315393.3	15.0000	0.0000
18.5000	273473.0	18.5000	-0.0000
23.9940	219920.8	23.9940	0.0000
29.0000	181414.8	29.0000	-0.0000
32.5000	159100.0	32.5000	0.0000

Temperature ITS-90 =  $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$  (°C)

Residual = instrument temperature - bath temperature

n = instrument output



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9288  
 CALIBRATION DATE: 26-Feb-15

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -9.830211e-001  
 h = 1.509796e-001  
 i = -8.585437e-005  
 j = 2.874494e-005

CPcor = -9.5700e-008  
 CTcor = 3.2500e-006  
 WBOTC = 9.7955e-008


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2551.92	0.00000	0.00000
1.0000	34.6602	2.96389	5107.67	2.96390	0.00001
4.5000	34.6404	3.26976	5301.18	3.26975	-0.00001
15.0000	34.5971	4.24756	5876.56	4.24756	0.00000
18.5000	34.5868	4.59120	6065.59	4.59120	0.00001
23.9940	34.5717	5.14563	6358.46	5.14562	-0.00001
29.0000	34.5650	5.66577	6621.20	5.66578	0.00001
32.5000	34.5600	6.03633	6802.02	6.03633	-0.00000

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9288  
CALIBRATION DATE: 24-Feb-15

Slocum Payload CTD PRESSURE CALIBRATION DATA  
FSR: 1450 psia S/N 4326844

**COEFFICIENTS:**

PA0 = -7.571650e-002	PTCA0 = 5.245973e+005
PA1 = 4.544461e-003	PTCA1 = -2.377949e+000
PA2 = -2.023587e-011	PTCA2 = 7.345523e-002
PTEMPA0 = 1.476112e+002	PTCB0 = 2.511588e+001
PTEMPA1 = -6.604377e-002	PTCB1 = -2.250000e-004
PTEMPA2 = 2.712419e-007	PTCB2 = 0.000000e+000

**PRESSURE SPAN CALIBRATION**

**THERMAL CORRECTION**

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS	TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
14.78	527867.0	1919.0	14.86	0.01	32.50	1756	527926.25
315.03	593919.0	1917.0	314.99	-0.00	29.00	1809	527922.00
615.01	659976.0	1916.0	614.97	-0.00	23.99	1886	527911.00
914.99	726087.0	1915.0	915.02	0.00	18.50	1971	527898.75
1214.92	792233.0	1914.0	1215.05	0.01	15.00	2025	527913.25
1464.89	847305.0	1910.0	1464.72	-0.01	4.50	2187	527922.50
1214.88	792229.0	1913.0	1215.03	0.01	1.00	2241	527919.75
915.04	726084.0	1914.0	915.01	-0.00			
615.04	659981.0	1915.0	615.00	-0.00			
315.08	593919.0	1917.0	314.99	-0.01			
14.78	527853.0	1918.0	14.80	0.00			

	TEMP (ITS90)	SPAN (mV)
	-5.00	25.12
	35.00	25.11

$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$   
 $x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$   
 $n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$   
 $\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$

