



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9057  
 CALIBRATION DATE: 12-Jan-18

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.063559e-004  
 a1 = 3.086859e-004  
 a2 = -4.685448e-006  
 a3 = 2.048573e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	587734.1	1.0000	0.0000
4.5000	502134.1	4.5000	-0.0000
15.0000	319398.2	15.0000	0.0000
18.5000	276421.6	18.5000	0.0000
24.0000	221600.7	24.0000	-0.0000
29.0000	182380.0	28.9999	-0.0001
32.5000	159669.7	32.5000	0.0000

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{a_0 + a_1 [\ln(n)] + a_2 [\ln^2(n)] + a_3 [\ln^3(n)]\} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9057  
 CALIBRATION DATE: 12-Jan-18


Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.829348e-001                      CPcor = -9.5700e-008  
 h = 1.326842e-001                      CTcor = 3.2500e-006  
 i = -2.946968e-004                      WBOTC = 1.4056e-006  
 j = 3.846522e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2727.06	0.00000	0.00000
1.0000	34.6897	2.96617	5465.08	2.96617	-0.00000
4.5000	34.6683	3.27213	5672.26	3.27213	-0.00000
15.0000	34.6262	4.25075	6288.44	4.25076	0.00001
18.5000	34.6175	4.59483	6490.90	4.59484	0.00001
24.0000	34.6081	5.15107	6805.12	5.15106	-0.00001
29.0000	34.6031	5.67132	7086.06	5.67130	-0.00002
32.5000	34.6001	6.04254	7279.70	6.04255	0.00001

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$   
 $t = \text{temperature (°C)}$ ;  $p = \text{pressure (decibars)}$ ;  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;  
 $\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$   
 $\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9057  
 CALIBRATION DATE: 08-Jan-18

Slocum Payload CTD PRESSURE CALIBRATION DATA  
 1450 psia S/N 3791765

COEFFICIENTS:

PA0 =	1.253453e-001	PTCA0 =	5.250186e+005
PA1 =	4.560399e-003	PTCA1 =	1.424170e+000
PA2 =	-1.228350e-011	PTCA2 =	-8.210813e-003
PTEMPA0 =	-6.868934e+001	PTCB0 =	2.533450e+001
PTEMPA1 =	5.179394e-002	PTCB1 =	-1.000000e-004
PTEMPA2 =	-3.989164e-007	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.53	528206.5	1773.1	14.54	0.00	32.50	1984	528265.90
301.44	591115.6	1774.4	301.40	-0.00	29.00	1914	528267.10
588.66	654128.6	1775.0	588.64	-0.00	24.00	1815	528260.90
875.88	717159.2	1775.7	875.86	-0.00	18.50	1706	528252.90
1163.12	780217.3	1776.3	1163.11	-0.00	15.00	1636	528249.50
1450.28	843274.0	1776.7	1450.25	-0.00	4.50	1429	528235.40
1163.19	780248.4	1776.3	1163.25	0.00	1.00	1360	528232.90
875.96	717185.1	1776.1	875.98	0.00			
588.68	654140.3	1776.0	588.69	0.00			
301.33	591098.2	1777.1	301.32	-0.00			
14.52	528206.1	1777.5	14.54	0.00			

	TEMPERATURE (°C)	SPAN
	-5.00	25.34
	35.00	25.33

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 08-Jan-18 0.00

