

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9285  
CALIBRATION DATE: 07-Mar-15

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

### COEFFICIENTS:


a0 = -8.274431e-005  
a1 = 3.088440e-004  
a2 = -4.856210e-006  
a3 = 2.098755e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0001	573079.4	1.0001	0.0000
4.5000	489112.0	4.4999	-0.0001
15.0000	310175.4	15.0001	0.0001
18.5000	268178.4	18.5000	0.0000
24.0000	214670.2	23.9998	-0.0002
29.0000	176436.8	29.0002	0.0002
32.5000	154326.8	32.4999	-0.0001

Temperature ITS-90 =  $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$  (°C)

Residual = instrument temperature - bath temperature

n = instrument output



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9285  
CALIBRATION DATE: 07-Mar-15

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -9.902783e-001  
h = 1.568274e-001  
i = -1.801139e-004  
j = 3.765629e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 1.3255e-007

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2514.57	0.00000	0.00000
1.0001	34.6512	2.96320	5020.16	2.96321	0.00001
4.5000	34.6313	3.26898	5210.05	3.26898	-0.00001
15.0000	34.5883	4.24659	5774.77	4.24659	-0.00000
18.5000	34.5791	4.59029	5960.36	4.59028	-0.00000
24.0000	34.5689	5.14588	6248.48	5.14589	0.00001
29.0000	34.5636	5.66557	6506.14	5.66556	-0.00001
32.5000	34.5617	6.03659	6683.72	6.03629	-0.00030

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$


$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = instrument conductivity - bath conductivity

Date, Slope Correction

● 07-Mar-15 1.0000000



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9285  
CALIBRATION DATE: 24-Feb-15

Slocum Payload CTD PRESSURE CALIBRATION DATA  
FSR: 1450 psia S/N 4333816

COEFFICIENTS:

PA0 =	9.296916e-003	PTCA0 =	5.243658e+005
PA1 =	4.610702e-003	PTCA1 =	-1.195399e+000
PA2 =	-2.417621e-011	PTCA2 =	5.823210e-002
PTEMPA0 =	1.361930e+002	PTCB0 =	2.504050e+001
PTEMPA1 =	-6.412924e-002	PTCB1 =	-1.000000e-004
PTEMPA2 =	1.678661e-008	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS	TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
14.78	527589.0	1783.0	14.86	0.01	32.50	1618	527650.80
315.03	592705.0	1781.0	315.01	-0.00	29.00	1672	527643.80
615.01	657829.0	1779.0	614.98	-0.00	24.00	1750	527625.80
914.99	723012.0	1779.0	915.03	0.00	18.50	1836	527632.80
1214.92	788233.0	1776.0	1215.04	0.01	15.00	1891	527624.60
1464.89	842551.0	1774.0	1464.74	-0.01	4.50	2055	527615.00
1214.88	788227.0	1775.0	1215.01	0.01	1.00	2109	527632.40
915.04	723011.0	1776.0	915.02	-0.00			
615.04	657830.0	1777.0	614.99	-0.00	TEMP (ITS90)	SPAN (mV)	
315.08	592703.0	1779.0	315.00	-0.01	-5.00	25.04	
14.78	527573.0	1779.0	14.79	0.00	35.00	25.04	

$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$   
 $x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$   
 $n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$   
 $\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$

Date, Avg Delta P %FS

● 24-Feb-15 0.00

