



**SEA-BIRD
SCIENTIFIC**

SEA-BIRD ELECTRONICS, INC.
13431 NE 20th Street
Bellevue, Washington 98005 USA

Phone +1-425-643-9866
Fax +1-425-643-9954
www.seabird.com

SERVICE REPORT

**Service Request
Date**

**1005501447
25-MAR-2017**

PRODUCT INFORMATION

Item: SLOCUM.LEGACY
Item Description: (LEGACY) Slocum Glider
Serial: 9056

Special Notes

Services Requested:
Evaluate/Repair Instrumentation.
Perform Routine Calibration Service.

Problems Found:
No problems found

Services Performed:
Perform initial diagnostic evaluation.
Performed "POST" cruise calibration.
Performed pressure calibration.
Performed complete system check and full diagnostic evaluation.
Installed NEW AF24173 Anti-foulant cylinder(s).

Item	Item Description	Qty
CAL_SLOCUM	CALIBRATE SLOCUM CONDUCTIVITY AND TEMPERATURE SENSORS (FRRF)	1
CNCRTSLOCUM	CONFIRM & RE-CERTIFY WEBB SLOCUM GLIDER CTD (FRRF)	1
REPLACEAF	EXTRA CHARGE TO INSTALL ONE ANTIFOULANT DEVICE, INCLUDES ONE 801542.1. (FRRF)	1
PCAL_SLOCUM	CALIBRATE SLOCUM PRESSURE SENSOR (FRRF)	1

Unbilled Items

Item	Item Description	Qty
801542.1	AF24173 ANTI-FOULANT, SINGLE CYLINDER, V2	1

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9056
 CALIBRATION DATE: 26-Feb-17

Slocum Payload CTD TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.093713e-004
 a1 = 3.088652e-004
 a2 = -4.613861e-006
 a3 = 2.032910e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	571313.8	0.9999	-0.0001
4.5000	488373.4	4.5002	0.0002
15.0000	311166.4	15.0000	-0.0000
18.4999	269447.0	18.4996	-0.0003
24.0000	216181.8	24.0004	0.0004
29.0000	178054.0	29.0000	-0.0000
32.5000	155962.6	32.4999	-0.0001

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9056
CALIBRATION DATE: 26-Feb-17

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.850771e-001
h = 1.365074e-001
i = -5.065597e-004
j = 5.490522e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 2.4111e-006


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2695.81	0.00000	0.00000
1.0000	34.7492	2.97078	5405.70	2.97076	-0.00002
4.5000	34.7297	3.27736	5610.88	3.27738	0.00003
15.0000	34.6876	4.25749	6220.62	4.25749	0.00000
18.4999	34.6792	4.60213	6420.94	4.60211	-0.00002
24.0000	34.6702	5.15929	6731.84	5.15929	0.00000
29.0000	34.6660	5.68047	7009.77	5.68049	0.00002
32.5000	34.6642	6.05246	7201.31	6.05245	-0.00001

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9056

CALIBRATION DATE: 22-Feb-17

Slocum Payload CTD PRESSURE CALIBRATION DATA

1450 psia S/N 3806514

COEFFICIENTS:

PA0 = -1.807734e-001

PA1 = 4.614206e-003

PA2 = -1.849638e-011

PTEMPA0 = -6.901078e+001

PTEMPA1 = 5.183566e-002

PTEMPA2 = -4.692761e-007

PTCA0 = 5.245200e+005

PTCA1 = -1.886116e+000

PTCA2 = 8.101812e-004

PTCB0 = 2.521537e+001

PTCB1 = 8.750000e-004

PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.70	527707.0	1774.0	14.70	-0.00	32.50	1995	527876.20
314.97	592844.0	1774.0	314.94	-0.00	29.00	1924	527888.60
614.96	657954.0	1774.0	614.91	-0.00	24.00	1824	527898.80
914.93	723106.0	1775.0	914.91	-0.00	18.50	1715	527904.00
1214.95	788300.0	1774.0	1214.95	0.00	15.00	1645	527908.20
1464.92	842637.0	1774.0	1464.90	-0.00	4.50	1437	527932.80
1214.90	788294.0	1774.0	1214.92	0.00	1.00	1367	527937.80
914.93	723120.0	1773.0	914.98	0.00			
614.94	657965.0	1773.0	614.96	0.00			
314.93	592849.0	1773.0	314.97	0.00			
14.70	527709.0	1772.0	14.71	0.00			

THERMAL CORRECTION

TEMPERATURE (°C)	SPAN (mV)
-5.00	25.21
35.00	25.25

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 22-Feb-17 0.00

