

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9235  
 CALIBRATION DATE: 17-Sep-14

Slocum Payload CTD PRESSURE CALIBRATION DATA  
 FSR: 1450 psia S/N 4233472

**COEFFICIENTS:**

PA0 =	1.559447e-001	PTCA0 =	5.232477e+005
PA1 =	4.531086e-003	PTCA1 =	3.250209e+000
PA2 =	-1.284814e-011	PTCA2 =	-7.782236e-002
PTEMPA0 =	-7.414059e+001	PTCB0 =	2.534013e+001
PTEMPA1 =	4.963390e-002	PTCB1 =	-1.750000e-004
PTEMPA2 =	-3.433862e-007	PTCB2 =	0.000000e+000

**PRESSURE SPAN CALIBRATION**


**THERMAL CORRECTION**

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS	TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
14.50	526459.0	1984.0	14.56	0.00	32.50	2182	526498.75
314.79	592707.0	1990.0	314.72	-0.00	29.00	2109	526508.25
614.74	658945.0	1990.0	614.72	-0.00	24.00	2005	526514.00
914.80	725230.0	1992.0	914.83	0.00	18.50	1891	526510.00
1214.74	791494.0	1992.0	1214.72	-0.00	15.00	1819	526505.75
1464.73	846742.0	1992.0	1464.68	-0.00	4.50	1602	526492.75
1214.73	791509.0	1992.0	1214.79	0.00	1.00	1530	526480.00
914.74	725223.0	1992.0	914.80	0.00			
614.80	658959.0	1992.0	614.79	-0.00			
314.82	592719.0	1992.0	314.77	-0.00			
14.50	526448.0	1993.0	14.51	0.00			

TEMP (ITS90)	SPAN (mV)
-5.00	25.34
35.00	25.33

$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$   
 $x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$   
 $n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$   
 $\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9235  
CALIBRATION DATE: 23-Sep-14

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## COEFFICIENTS:


a0 = -1.367772e-004  
a1 = 3.146117e-004  
a2 = -5.065285e-006  
a3 = 2.180317e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	564038.8	1.0001	0.0001
4.5000	482535.8	4.4999	-0.0001
15.0000	308095.3	15.0001	0.0001
18.5000	266958.0	18.5001	0.0001
23.9999	214406.5	23.9997	-0.0002
29.0000	176737.5	29.0000	0.0000
32.5000	154898.0	32.5000	0.0000

Temperature ITS-90 =  $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$  (°C)

Residual = instrument temperature - bath temperature

n = instrument output



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9235  
 CALIBRATION DATE: 23-Sep-14

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -9.877144e-001	CPcor = -9.5700e-008
h = 1.523432e-001	CTcor = 3.2500e-006
i = -2.379939e-004	WBOTC = 7.5185e-007
j = 4.083202e-005	

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2549.11	0.00000	0.00000
1.0000	34.7066	2.96748	5097.88	2.96748	0.00000
4.5000	34.6868	3.27371	5290.94	3.27371	-0.00000
15.0000	34.6444	4.25275	5864.99	4.25275	0.00000
18.5000	34.6349	4.59689	6053.60	4.59690	0.00001
23.9999	34.6243	5.15320	6346.35	5.15320	-0.00000
29.0000	34.6183	5.67353	6608.14	5.67352	-0.00001
32.5000	34.6139	6.04467	6788.53	6.04468	0.00000

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperatur e[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = instrument conductivity - bath conductivity

