



SEA-BIRD

**SEA-BIRD ELECTRONICS, INC.**

13431 NE 20<sup>th</sup> Street  
Bellevue, Washington 98005 USA

Phone +1-425-643-9866  
Fax +1-425-643-9954  
www.seabird.com

**SERVICE REPORT**

**Service Request  
Date**

**1005501814  
06-MAY-2017**

**PRODUCT INFORMATION**

**Item:** SLOCUM.LEGACY  
**Item Description:** (LEGACY) Slocum Glider  
**Serial:** 9085

**Special Notes**

Services Requested:  
Evaluate/Repair Instrumentation.  
Perform Routine Calibration Service.  
Replace Antifoulant Device(s).

Services Performed:

Perform initial diagnostic evaluation.  
Performed pressure calibration.  
Performed "POST" cruise calibration.  
Installed NEW AF24173 Anti-foulant cylinder(s).

Item	Item Description	Qty
CAL_SLOCUM	CALIBRATE SLOCUM CONDUCTIVITY AND TEMPERATURE SENSORS (FRRF)	1
CNCRTSLOCUM	CONFIRM & RE-CERTIFY WEBB SLOCUM GLIDER CTD (FRRF)	1
REPLACEAF	EXTRA CHARGE TO INSTALL ONE ANTIFOULANT DEVICE, INCLUDES ONE 801542.1. (FRRF)	1

**Unbilled Items**

Item	Item Description	Qty
801542.1	AF24173 ANTI-FOULANT, SINGLE CYLINDER, V2	1

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9085  
CALIBRATION DATE: 28-Apr-17

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

**COEFFICIENTS:**


a0 = -3.852659e-005  
a1 = 2.934099e-004  
a2 = -3.460815e-006  
a3 = 1.749545e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	566317.4	1.0000	0.0000
4.5000	484141.6	4.4999	-0.0001
15.0000	308505.8	15.0000	0.0000
18.5000	267144.8	18.5001	0.0001
24.0000	214344.6	24.0000	0.0000
29.0000	176536.0	28.9999	-0.0001
32.5000	154626.4	32.5001	0.0001

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$$

$$\text{Residual (°C)} = \text{instrument temperature} - \text{bath temperature}$$



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9085  
CALIBRATION DATE: 28-Apr-17

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter


**COEFFICIENTS:**

g = -9.906300e-001  
h = 1.367170e-001  
i = -2.738901e-004  
j = 3.950898e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 3.0905e-007

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2696.26	0.00000	0.00000
1.0000	34.7294	2.96924	5388.31	2.96924	-0.00000
4.5000	34.7100	3.27568	5592.27	3.27568	-0.00000
15.0000	34.6684	4.25539	6198.65	4.25540	0.00001
18.5000	34.6595	4.59981	6397.87	4.59980	-0.00001
24.0000	34.6497	5.15658	6707.10	5.15657	-0.00000
29.0000	34.6438	5.67724	6983.54	5.67724	0.00000
32.5000	34.6395	6.04864	7173.47	6.04760	-0.00104

f = Instrument Output(Hz) \* sqrt(1.0 + WBOTC \* t) / 1000.0  
t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;  
Conductivity (S/m) = (g + h \* f<sup>2</sup> + i \* f<sup>3</sup> + j \* f<sup>4</sup>) / 10 (1 + δ \* t + ε \* p)  
Residual (Siemens/meter) = instrument conductivity - bath conductivity



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9085  
CALIBRATION DATE: 24-Apr-17

Slocum Payload CTD PRESSURE CALIBRATION DATA  
1450 psia S/N 3846544

**COEFFICIENTS:**

PA0 = 1.027209e-001	PTCA0 = 5.246226e+005
PA1 = 4.493376e-003	PTCA1 = 6.052008e-001
PA2 = -2.049231e-011	PTCA2 = 1.561417e-002
PTEMPA0 = -7.320175e+001	PTCB0 = 2.545600e+001
PTEMPA1 = 5.141549e-002	PTCB1 = -6.000000e-004
PTEMPA2 = -5.890481e-007	PTCB2 = 0.000000e+000

**PRESSURE SPAN CALIBRATION**

**THERMAL CORRECTION**

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.47	527864.7	1916.6	14.58	0.01	32.50	2107	527933.00
301.48	591681.1	1917.5	301.39	-0.01	29.00	2035	527930.00
588.74	655646.1	1918.5	588.71	-0.00	24.00	1934	527923.80
875.95	719642.1	1918.8	876.00	0.00	18.50	1821	527913.20
1163.18	783659.7	1919.1	1163.22	0.00	15.00	1751	527909.60
1450.34	847677.6	1919.4	1450.27	-0.00	4.50	1538	527902.60
1163.17	783660.7	1919.0	1163.22	0.00	1.00	1468	527898.20
875.90	719628.7	1918.9	875.94	0.00			
588.74	655645.2	1918.6	588.71	-0.00	TEMPERATURE (°C)	SPAN (mV)	
301.47	591688.0	1918.3	301.42	-0.00	-5.00	25.46	
14.47	527834.9	1918.1	14.44	-0.00	35.00	25.43	

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 24-Apr-17 -0.00

