



SEA-BIRD
SCIENTIFIC

Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone
Fax

+1-425-643-9866
+1-425-643-9954
www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005507783
28-JUN-2019
315621426

CUSTOMER INFORMATION

Name: TELEDYNE WEBB RESEARCH
Account : 40280819
CHARLES STILL
CHARLES.STILL@TELEDYNE.COM
508-563-1000

PO Number:
208512

Bill To Address

ATTN: ACCOUNTS PAYABLE;1026 N. Williamson Blvd.;
Daytona Beach,FL,32114,US

Ship To Address

BUSINESS UNIT OF TELEDYNE INSTRUMENT INC;49
EDGERTON DRIVE;
NORTH FALMOUTH,MA,02556,US

PRODUCT INFORMATION

Item: SLOCUM.LEGACY
Item Description: (LEGACY) Slocum Glider
Serial: 9023

Special Notes

Services Requested:
Evaluate/Repair Instrumentation.
Perform Routine Calibration Service.

Problems Found:

The hardcoating on the endcap is getting thin/corroded. It is recommended to replace endcap during next service.

Services Performed:

Performed initial diagnostic evaluation.
Performed "POST" cruise calibration.
Performed pressure calibration.
Installed NEW AF24173 Anti-foulant cylinder.

| Item | Item Description | Qty |
|-------------|--|-----|
| CAL_SLOCUM | Calibrate SLOCUM conductivity and temperature sensors | 1 |
| CNCRSLOCUM | Confirm & Re-certify Webb SLOCUM Glider CTD | 1 |
| REPLACEAF | Extra charge to install one antifoulant device, includes one 801542.1. | 1 |
| PCAL_SLOCUM | Calibrate SLOCUM pressure sensor | 1 |

Unbilled Items

| Item | Item Description | Qty |
|----------|---|-----|
| 801542.1 | AF24173 ANTI-FOULANT, SINGLE CYLINDER, V2 | 1 |



SEA-BIRD
SCIENTIFIC

Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9023
CALIBRATION DATE: 31-May-19

Slocum Payload CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.080157e-004
a1 = 3.078367e-004
a2 = -4.543024e-006
a3 = 2.028552e-007

| BATH TEMP (° C) | INSTRUMENT OUTPUT (counts) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| 1.0000 | 572952.8 | 1.0000 | 0.0000 |
| 4.5000 | 489950.3 | 4.4999 | -0.0001 |
| 15.0000 | 312457.7 | 15.0000 | -0.0000 |
| 18.5000 | 270636.2 | 18.5001 | 0.0001 |
| 24.0000 | 217235.4 | 24.0000 | 0.0000 |
| 29.0000 | 178985.4 | 28.9998 | -0.0002 |
| 32.4999 | 156813.8 | 32.5000 | 0.0001 |

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9023
 CALIBRATION DATE: 31-May-19


Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.936690e-001 CPcor = -9.5700e-008
 h = 1.401546e-001 CTcor = 3.2500e-006
 i = -2.217543e-004 WBOTC = -3.4385e-007
 j = 3.600398e-005

| BATH TEMP (° C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (Hz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|--------------------|-------------------|--------------------|---------------------------|--------------------------|-------------------|
| 22.0000 | 0.0000 | 0.00000 | 2665.87 | 0.00000 | 0.00000 |
| 1.0000 | 34.6514 | 2.96321 | 5316.48 | 2.96322 | 0.00001 |
| 4.5000 | 34.6318 | 3.26903 | 5517.45 | 3.26902 | -0.00001 |
| 15.0000 | 34.5893 | 4.24670 | 6115.08 | 4.24670 | -0.00001 |
| 18.5000 | 34.5802 | 4.59042 | 6311.48 | 4.59042 | 0.00000 |
| 24.0000 | 34.5692 | 5.14592 | 6616.27 | 5.14592 | 0.00001 |
| 29.0000 | 34.5596 | 5.66499 | 6888.57 | 5.66498 | -0.00000 |
| 32.4999 | 34.5496 | 6.03471 | 7075.76 | 6.03433 | -0.00037 |

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0
 t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;
 Conductivity (S/m) = (g + h * f² + i * f³ + j * f⁴) / (1 + δ * t + ε * p)
 Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9023
 CALIBRATION DATE: 22-May-19

Slocum Payload CTD PRESSURE CALIBRATION DATA
 1450 psia S/N 3647167

COEFFICIENTS:

| | | | |
|-----------|----------------|---------|----------------|
| PA0 = | -5.320865e-002 | PTCA0 = | 5.236884e+005 |
| PA1 = | 4.569282e-003 | PTCA1 = | -8.642211e-001 |
| PA2 = | -1.642226e-011 | PTCA2 = | 1.518158e-002 |
| PTEMPA0 = | -6.910766e+001 | PTCB0 = | 2.539625e+001 |
| PTEMPA1 = | 5.235000e-002 | PTCB1 = | -3.500000e-004 |
| PTEMPA2 = | -6.021189e-007 | PTCB2 = | 0.000000e+000 |

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

| PRESSURE (PSIA) | INSTRUMENT OUTPUT (counts) | THERMISTOR OUTPUT (volts) | COMPUTED PRESSURE (PSIA) | RESIDUAL (%FSR) | TEMP (°C) | THERMISTOR OUTPUT (volts) | INSTRUMENT OUTPUT (counts) |
|-----------------|----------------------------|---------------------------|--------------------------|-----------------|------------------|---------------------------|----------------------------|
| 14.59 | 526877.6 | 1778.1 | 14.58 | -0.00 | 32.50 | 1986 | 526922.70 |
| 300.91 | 589534.5 | 1781.8 | 300.89 | -0.00 | 29.00 | 1916 | 526925.20 |
| 588.17 | 652416.2 | 1782.9 | 588.10 | -0.00 | 24.00 | 1817 | 526925.90 |
| 875.46 | 715343.1 | 1782.6 | 875.39 | -0.00 | 18.50 | 1707 | 526924.20 |
| 1162.67 | 778286.2 | 1782.5 | 1162.62 | -0.00 | 15.00 | 1637 | 526925.70 |
| 1450.01 | 841294.3 | 1782.1 | 1450.02 | 0.00 | 4.50 | 1430 | 526933.10 |
| 1162.73 | 778323.4 | 1781.2 | 1162.79 | 0.00 | 1.00 | 1361 | 526935.30 |
| 875.47 | 715368.4 | 1780.7 | 875.50 | 0.00 | | | |
| 588.07 | 652411.0 | 1779.8 | 588.08 | 0.00 | TEMPERATURE (°C) | SPAN | |
| 300.95 | 589577.0 | 1778.2 | 301.08 | 0.01 | -5.00 | 25.40 | |
| 14.59 | 526873.6 | 1778.3 | 14.56 | -0.00 | 35.00 | 25.38 | |

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 22-May-19 -0.00

