

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9015
CALIBRATION DATE: 23-Dec-14

Slocum Payload CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.292092e-004
a1 = 3.119697e-004
a2 = -4.838602e-006
a3 = 2.124698e-007

| BATH TEMP (ITS-90) | INSTRUMENT OUTPUT | INST TEMP (ITS-90) | RESIDUAL (ITS-90) |
|-----------------------|----------------------|-----------------------|----------------------|
| 1.0000 | 565154.6 | 1.0000 | -0.0000 |
| 4.5000 | 483599.6 | 4.5001 | 0.0001 |
| 15.0000 | 308989.8 | 14.9999 | -0.0001 |
| 18.5000 | 267789.3 | 18.5001 | 0.0001 |
| 24.0000 | 215142.7 | 23.9999 | -0.0001 |
| 29.0000 | 177396.2 | 29.0001 | 0.0001 |
| 32.5000 | 155505.6 | 32.5000 | -0.0000 |

Temperature ITS-90 = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature

n = instrument output



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9015
CALIBRATION DATE: 19-Dec-14

Slocum Payload CTD PRESSURE CALIBRATION DATA
FSR: 1450 psia S/N 3626944

COEFFICIENTS:

| | | | |
|-----------|----------------|---------|----------------|
| PA0 = | 5.969189e-001 | PTCA0 = | 5.248687e+005 |
| PA1 = | 4.855290e-003 | PTCA1 = | 1.007912e+001 |
| PA2 = | -1.853051e-011 | PTCA2 = | -2.049137e-001 |
| PTEMPA0 = | -6.962622e+001 | PTCB0 = | 2.519613e+001 |
| PTEMPA1 = | 5.157931e-002 | PTCB1 = | 2.250000e-004 |
| PTEMPA2 = | -4.604487e-007 | PTCB2 = | 0.000000e+000 |


PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

| PRESSURE PSIA | INST OUTPUT | THERMISTOR OUTPUT | COMPUTED PRESSURE | ERROR %FS | TEMP ITS90 | THERMISTOR OUTPUT | INST OUTPUT |
|------------------|----------------|----------------------|----------------------|--------------|---------------|----------------------|----------------|
| 14.56 | 527876.0 | 1824.0 | 14.60 | 0.00 | 32.50 | 2016 | 527936.70 |
| 314.56 | 589753.0 | 1825.0 | 314.89 | 0.02 | 29.00 | 1946 | 527954.50 |
| 614.94 | 651600.0 | 1827.0 | 614.89 | -0.00 | 24.00 | 1846 | 527957.60 |
| 914.95 | 713481.0 | 1827.0 | 914.92 | -0.00 | 18.50 | 1735 | 527945.80 |
| 1214.92 | 775389.0 | 1828.0 | 1214.93 | 0.00 | 15.00 | 1665 | 527930.80 |
| 1464.94 | 826991.0 | 1828.0 | 1464.90 | -0.00 | 4.50 | 1456 | 527871.70 |
| 1214.92 | 775396.0 | 1829.0 | 1214.97 | 0.00 | 1.00 | 1387 | 527840.70 |
| 914.95 | 713498.0 | 1828.0 | 915.00 | 0.00 | | | |
| 614.95 | 651611.0 | 1828.0 | 614.94 | -0.00 | | | |
| 314.90 | 589756.0 | 1828.0 | 314.90 | 0.00 | | | |
| 14.56 | 527863.0 | 1827.0 | 14.53 | -0.00 | | | |

| TEMP (ITS90) | SPAN (mV) |
|--------------|-----------|
| -5.00 | 25.20 |
| 35.00 | 25.20 |

$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$
 $x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$
 $n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$
 $\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9015
CALIBRATION DATE: 23-Dec-14

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.878740e-001
h = 1.354520e-001
i = -1.668583e-004
j = 2.985470e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -1.5415e-007

| BATH TEMP (ITS-90) | BATH SAL (PSU) | BATH COND (Siemens/m) | INST FREQ (Hz) | INST COND (Siemens/m) | RESIDUAL (Siemens/m) |
|-----------------------|-------------------|--------------------------|-------------------|--------------------------|-------------------------|
| 22.0000 | 0.0000 | 0.00000 | 2702.92 | 0.00000 | 0.00000 |
| 1.0000 | 34.7014 | 2.96708 | 5404.14 | 2.96709 | 0.00001 |
| 4.5000 | 34.6816 | 3.27327 | 5608.76 | 3.27326 | -0.00001 |
| 15.0000 | 34.6385 | 4.25210 | 6217.19 | 4.25210 | -0.00001 |
| 18.5000 | 34.6293 | 4.59623 | 6417.14 | 4.59624 | 0.00001 |
| 24.0000 | 34.6193 | 5.15255 | 6727.52 | 5.15255 | 0.00000 |
| 29.0000 | 34.6137 | 5.67286 | 7005.07 | 5.67286 | 0.00000 |
| 32.5000 | 34.6100 | 6.04407 | 7196.36 | 6.04407 | -0.00000 |

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperatur e[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

