



SEA-BIRD

Sea-Bird Electronics, Inc.  
13431 NE 20th Street  
Bellevue, WA 98005 United States

Phone  
Fax

+1-425-643-9866  
+1-425-643-9954  
www.seabird.com  
**1005508829**  
11-OCT-2019  
315749641

**SERVICE REPORT**

**Service Request**  
**Date**  
**Sales Order**

**CUSTOMER INFORMATION**

Name: TELEDYNE WEBB RESEARCH  
Account : 40280819  
CHUCK STILL  
CHARLES.STILL@TELEDYNE.COM  
508-563-1000

PO Number:

**Bill To Address**

ATTN: ACCOUNTS PAYABLE;1026 N. Williamson Blvd.;  
Daytona Beach,FL,32114,US

**Ship To Address**

BUSINESS UNIT OF TELEDYNE INSTRUMENT INC;49  
EDGERTON DRIVE;  
NORTH FALMOUTH,MA,02556,US

**PRODUCT INFORMATION**

Item: SLOCUM.50  
Item Description: SLOCUM GLIDER CTD, 1000 dBar, DIRECT GROUND  
Serial: 712-9362

**Special Notes**

Services Requested:  
Evaluate/Repair Instrumentation.  
Perform Routine Calibration Service.

Services Performed:

Perform initial diagnostic evaluation.  
Performed pressure calibration.  
Performed "POST" cruise calibration.  
Replaced the lithium back-up battery(s).  
Installed NEW AF24173 Anti-foulant cylinder(s).

Item	Item Description	Qty
CAL_SLOCUM	Calibrate SLOCUM conductivity and temperature sensors	1
CNCRTSLOCUM	Confirm & Re-certify Webb SLOCUM Glider CTD	1
REPLACEAF	Extra charge to install one antifoulant device, includes one 801542.1.	1
PCAL_SLOCUM	Calibrate SLOCUM pressure sensor	1

**Unbilled Items**

Item	Item Description	Qty
801542.1	AF24173 ANTI-FOULANT, SINGLE CYLINDER, V2	1
22096	LITHIUM COIN BATTERY, WITH TABS, BR1632A/HA	1



**SEA-BIRD**  
SCIENTIFIC

Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 9362  
CALIBRATION DATE: 27-Sep-19

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

**COEFFICIENTS:**


a0 = -1.451553e-004  
a1 = 3.096219e-004  
a2 = -4.380586e-006  
a3 = 1.991815e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	560568.8	1.0000	0.0000
4.5000	480368.4	4.5000	-0.0000
15.0000	308232.0	14.9999	-0.0001
18.4999	267506.4	18.5000	0.0001
24.0000	215375.8	24.0002	0.0002
29.0001	177933.4	28.9998	-0.0003
32.5000	156181.0	32.5001	0.0001

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





**SEA-BIRD**  
SCIENTIFIC

Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 9362  
CALIBRATION DATE: 27-Sep-19

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**


g = -9.900803e-001	CPCor = -9.5700e-008
h = 1.474709e-001	CTcor = 3.2500e-006
i = -1.997951e-004	WBOTC = 4.4980e-007
j = 3.556832e-005	

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2593.53	0.00000	0.00000
1.0000	34.7320	2.96945	5183.07	2.96945	0.00001
4.5000	34.7128	3.27592	5379.30	3.27592	-0.00000
15.0000	34.6714	4.25571	5962.81	4.25571	-0.00001
18.4999	34.6627	4.60018	6154.57	4.60017	-0.00001
24.0000	34.6526	5.15696	6452.24	5.15699	0.00003
29.0001	34.6460	5.67757	6718.32	5.67756	-0.00000
32.5000	34.6405	6.04879	6901.62	6.04879	-0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$   
 $t = \text{temperature } (^\circ\text{C}); p = \text{pressure (decibars); } \delta = \text{CTcor}; \epsilon = \text{CPCor};$   
 $\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$   
 $\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

Date, Slope Correction

- 13-Dec-17 0.9999773
- ▲ 27-Sep-19 1.0000000



**POST CRUISE  
CALIBRATION**



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9362  
 CALIBRATION DATE: 24-Sep-19

Slocum Payload CTD PRESSURE CALIBRATION DATA  
 1450 psia S/N 10087783

COEFFICIENTS:

PA0 =	4.861117e-002	PTCA0 =	5.258520e+005
PA1 =	4.519057e-003	PTCA1 =	4.139730e+000
PA2 =	-2.678397e-011	PTCA2 =	-1.464059e-001
PTEMPA0 =	1.317435e+002	PTCB0 =	2.510788e+001
PTEMPA1 =	-7.134121e-002	PTCB1 =	-2.500000e-005
PTEMPA2 =	-2.583941e-007	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.62	529098.0	1483.0	14.67	0.00	32.50	1384	529073.60
314.86	595527.0	1483.0	314.74	-0.01	29.00	1433	529090.00
614.87	662021.0	1482.0	614.88	0.00	24.00	1502	529102.40
914.84	728542.0	1480.0	914.89	0.00	18.50	1578	529120.00
1214.84	795106.0	1481.0	1214.87	0.00	15.00	1627	529122.80
1464.79	850593.0	1479.0	1464.74	-0.00	4.50	1772	529110.80
1214.82	795101.0	1482.0	1214.84	0.00	1.00	1821	529093.80
914.85	712051.0	1950.0	840.76	-5.11			
614.88	662025.0	1485.0	614.89	0.00	TEMPERATURE (°C)	SPAN	
314.91	595562.0	1486.0	314.90	-0.00	-5.00	25.11	
14.63	529095.0	1483.0	14.66	0.00	35.00	25.11	

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 24-Sep-19 -0.00

