

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9060  
CALIBRATION DATE: 08-Mar-13

SLOCUM PAYLOAD CTD  
PRESSURE CALIBRATION DATA  
1450 psia S/N 3805658

**COEFFICIENTS:**

PA0 = -1.266287e+000  
PA1 = 4.580788e-003  
PA2 = -1.463110e-011  
PTEMPA0 = -7.179874e+001  
PTEMPA1 = 5.141395e-002  
PTEMPA2 = -5.255467e-007

PTCA0 = 5.248416e+005  
PTCA1 = -1.204998e+001  
PTCA2 = -1.509298e-002  
PTCB0 = 2.532013e+001  
PTCB1 = 2.500000e-005  
PTCB2 = 0.000000e+000

**PRESSURE SPAN CALIBRATION**

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FSR
14.65	528047.0	1866.0	14.68	0.00
314.97	593594.0	1870.0	314.87	-0.01
614.96	659132.0	1870.0	614.89	-0.01
914.96	724709.0	1871.0	914.96	-0.00
1214.97	790302.0	1872.0	1214.97	-0.00
1465.04	844992.0	1873.0	1465.03	-0.00
1214.96	790297.0	1873.0	1214.95	-0.00
915.03	724731.0	1873.0	915.06	0.00
614.97	659180.0	1873.0	615.12	0.01
315.05	593617.0	1873.0	314.99	-0.00
14.65	528044.0	1873.0	14.69	0.00

**THERMAL CORRECTION**

TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
32.50	2073	527987.10
29.00	2001	528033.50
24.00	1900	528097.20
18.50	1789	528166.50
15.00	1719	528210.80
4.50	1507	528341.10
1.00	1437	528382.60
TEMP (ITS90)		SPAN (mV)
-5.00	25.32	
35.00	25.32	

$$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$


$$x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$$

Date, Avg Delta P %FS

● 08-Mar-13 0.00



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9060  
CALIBRATION DATE: 31-May-13

SLOCUM PAYLOAD CTD  
TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

a0 = -9.688467e-005  
a1 = 3.080132e-004  
a2 = -4.647427e-006  
a3 = 2.049510e-007


BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
0.9999	572953.0	0.9998	-0.0001
4.5000	489549.3	4.5002	0.0002
15.0000	311490.9	14.9999	-0.0001
18.5000	269600.7	18.5000	-0.0000
24.0000	216159.2	24.0000	0.0000
29.0000	177919.4	29.0002	0.0002
32.5000	155777.6	32.4999	-0.0001

Temperature ITS-90 =  $1 / \{a_0 + a_1 [\ln(n)] + a_2 [\ln^2(n)] + a_3 [\ln^3(n)]\} - 273.15$  (°C)

Residual = instrument temperature - bath temperature

Date, Delta T (mdeg C)

● 31-May-13 -0.00



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9060

CALIBRATION DATE: 31-May-13

SLOCUM PAYLOAD CTD

CONDUCTIVITY CALIBRATION DATA

PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -9.902037e-001

h = 1.323388e-001

i = -3.136565e-004

j = 3.986343e-005

CPcor = -9.5700e-008

CTcor = 3.2500e-006

WBOTC = 1.5807e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2741.15	0.00000	0.00000
0.9999	34.8131	2.97571	5485.10	2.97570	-0.00001
4.5000	34.7932	3.28276	5692.90	3.28277	0.00001
15.0000	34.7505	4.26439	6310.64	4.26439	-0.00001
18.5000	34.7413	4.60949	6513.60	4.60950	0.00000
24.0000	34.7312	5.16736	6828.60	5.16737	0.00000
29.0000	34.7256	5.68913	7110.22	5.68913	-0.00000

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = instrument conductivity - bath conductivity

Date, Slope Correction

● 31-May-13 1.0000000

